



EC Design-Examination Certificate

Directive 93/42/EEC on Medical Devices, Annex II Section 4

No. CE 633649
Issued To: **Internacional Farmacéutica, S.A. de C.V.**
Carreteraco No. 44
Colonia Parque San Andrés
Coyoacán
C.P. 04040
Ciudad de México
Mexico

In respect of:

Atramat® PGLA90 RAPID Polyglactin 910 Poly(glycolide-co-L-lactide) Quick Absorption Surgical Sutures

BSI has performed a design examination on the above devices in accordance with the Council Directive 93/42/EEC, Annex II Section 4. The design conforms to the requirements of this directive. For marketing of these products an additional Annex II excluding Section 4 certificate is required.

For and on behalf of BSI, a Notified Body for the above Directive (Notified Body Number 2797):

Gary E Slack, Senior Vice President Medical Devices

First Issued: **2015-11-30**

Date: **2020-05-15**

Expiry Date: **2024-05-26**

...making excellence a habit.™

Page 1 of 4

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.
This certificate was issued electronically and is bound by the conditions of the contract.

EC Design-Examination Certificate

Supplementary Information to CE 633649

Issued To:

Internacional Farmacéutica, S.A. de C.V.
Carreteraco No. 44
Colonia Parque San Andrés
Coyoacán
C.P. 04040
Ciudad de México
Mexico

Characteristics	Description	
Product Name	Atramat® PGLA90 RAPID Polyglactin 910 Poly(glycolide-co-L-lactide) Quick Absorption Surgical Sutures	
Needle Characteristics		
Needle Description	Premium, Fine, Heavy, Extra Heavy, Extra Heavy Loop, Premium Heavy, Premium Heavy Oval, Premium Regular, Premium Regular Oval, SUPRASHARP, ATRAGLIDE, SUPRAGLYDE, Supraglyde (Non Reflective), Regular.	
Needle Type & Codes	Reverse Cutting	C, CE, K, KE, P, PE, PM, P-P, P-PE.
	Spatula	E, EA, EE, E-E, E-EA, E-EE.
	Spatula Ultima	EP.
	Taper Point	F, FE, G, GE, GH, R, RE, S, SB, SBE, SE, SG, SGE, SGH, SJ, SR, SRE, S-SG, S-SR, SV, SVE, X, XE, X-Loop, T, E-S, E-SE, CVSB, CVSBE, CVSV, CVSVE.
	Taper Cutting	PW, PWE, PWH, KW, W, WE, YB, YBE, Y, YE, YJ, YV, YVE, CVYB, CVYBE, CVY, CVYE, CVYJ, CVYV, CVYVE.
	Conventional Cutting	P-Q, P-QE, N, NE, Q, QE, QH.
	Cutting	PU, HP, HPE, U.
	Without Needle	SS, SS-C, EZ-SS.
	Blunt Point:	Z, ZE.
Number of Needles	1 or 2.	

First Issued: **2015-11-30**

Date: **2020-05-15**

Expiry Date: **2024-05-26**

...making excellence a habit.™

EC Design-Examination Certificate

Supplementary Information to CE 633649

Issued To:

Internacional Farmacéutica, S.A. de C.V.
Carreteraco No. 44
Colonia Parque San Andrés
Coyoacán
C.P. 04040
Ciudad de México
Mexico

Characteristics	Description
Needle Attribute	Standard or Removable (PULL-OFF).
Needle Angle Range	1/4, 3/8, 5/8, 1/2, J, STRAIGHT and BI-CURVE.
Needle Length Range	6 mm (Min.) - 110 mm (Max.).
Suture Thread Characteristics	
Absorbable/Non-Absorbable	Absorbable.
Suture Color	Undyed.
Suture Gauge Sizes (USP)	6, 5, 4, 3, 2, 1, 0, 2-0, 3-0, 4-0, 5-0, 6-0, 7-0, 8-0.
Suture Length Range	15 cm (Min.) - 300 cm (Max.).
Quantity per Pouch	1-20.
Presentation	1 suture type or one family.
	Kit (one code or different codes, one family or different families).
	Box with 6, 12, 24 or 36 pouches in Premium or Conventional Line.

First Issued: **2015-11-30**

Date: **2020-05-15**

Expiry Date: **2024-05-26**

...making excellence a habit.™

Page 3 of 4

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.
 This certificate was issued electronically and is bound by the conditions of the contract.

EC Design-Examination Certificate

Supplementary Information to CE 633649

Issued To:

Internacional Farmacéutica, S.A. de C.V.
Carreteraco No. 44
Colonia Parque San Andrés
Coyoacán
C.P. 04040
Ciudad de México
Mexico

Certificate History

Date	Reference Number	Action
30 November 2015	10154701	First issue. Transfer from another Notified Body and Renewal.
25 February 2019	8588105	Traceable to NB 0086.
Current	3060859	Certificate Renewal. Administrative address change, Legal Manufacturer. Administrative address change, EU Rep. Addition of needle codes in the 'supplementary information' table of certificate.

First Issued: **2015-11-30**

Date: **2020-05-15**

Expiry Date: **2024-05-26**

...making excellence a habit.™

Page 4 of 4

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.
 This certificate was issued electronically and is bound by the conditions of the contract.