

EC Design-Examination Certificate

Directive 93/42/EEC on Medical Devices, Annex II Section 4

No.**CE 633673**

Issued To:

**Internacional Farmacéutica, S.A. de C.V.
Carreteraco No. 44
Colonia Parque San Andrés
Coyoacán
C.P. 04040
Ciudad de México
Mexico**

In respect of:

Atramat® Polyester Surgical Suture

BSI has performed a design examination on the above devices in accordance with the Council Directive 93/42/EEC, Annex II Section 4. The design conforms to the requirements of this directive. For marketing of these products an additional Annex II excluding Section 4 certificate is required.

For and on behalf of BSI, a Notified Body for the above Directive (Notified Body Number 2797):



Gary E Slack, Senior Vice President Medical Devices

First Issued: **2016-02-04**Date: **2020-05-05**Expiry Date: **2024-05-26**

...making excellence a habit.™

Page 1 of 4

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.
This certificate was issued electronically and is bound by the conditions of the contract.

EC Design-Examination Certificate

Supplementary Information to CE 633673

Issued To:

Internacional Farmacéutica, S.A. de C.V.
Carreteraco No. 44
Colonia Parque San Andrés
Coyoacán
C.P. 04040
Ciudad de México
Mexico

Characteristics	Description	
Product Name	Atramat® Polyester surgical suture	
Needle Characteristics		
Needle Description	Premium, Fine, Heavy, Extra Heavy, Premium Heavy, Premium Regular, SUPRASHARP , Regular, ATRAGLIDE , SUPRAGLYDE , SUPRAGLYDE (Non reflective)	
Needle Type & Codes	Reverse Cutting	C, CE, K, KE, P, PE, PM, P-P, P-PE.
	Spatula	E, EA, EE, E-E, E-EA, E-EE
	Taper Point	F, FE, G, GE, GH, R, RE, S, SB, SBE, SE, SG, SGE, SGH, SJ, SR, SRE, SV, SVE, X, XE, E-S, E-SE, CVSB, CVSBE, CVS, CVSVE
	Taper Cutting	KW, PW, TW, W, WE, Y, YB, YBE, YV, YVE, YE, YJ, CVY, CVYE, CVYB, CVYVE, CVYJ
	Conventional Cutting	N, NE, P-Q, P-QE, Q, QE
	Cutting	HP, HPE, PU, U
	Without Needle	SS, SS-C, EZ-SS
Number of Needles	1 or 2	
Needle Attribute	Standard or Removable (PULL-OFF)	
Needle Angle Range	1/4, 3/8, 5/8, 1/2, J, STRAIGHT and BI-CURVE	
Needle Length Range	7 mm (Min.) – 254 mm (Max.)	
Suture Thread Characteristics		
Absorbable/Non-Absorbable	Non-Absorbable	
Suture Color	Green or White	

First Issued: **2016-02-04**

Date: **2020-05-05**

Expiry Date: **2024-05-26**

...making excellence a habit.™

EC Design-Examination Certificate

Supplementary Information to CE 633673

Issued To:

**Internacional Farmacéutica, S.A. de C.V.
Carreteraco No. 44
Colonia Parque San Andrés
Coyoacán
C.P. 04040
Ciudad de México
Mexico**

Characteristics	Description
Suture Gauge Sizes (USP)	5, 4, 3, 2, 1, 0, 2-0, 3-0, 4-0, 5-0, 6-0, 7-0, 8-0
Suture Length Range	15 cm (Min.) - 300 cm (Max.)
Quantity per Pouch	1-15
Presentation	1 suture type or one family
	With or without pledget
	Kit (one code or different codes, one family or different families)
	Box with 6, 12, 24 or 36 pouches in Premium or Conventional Line

First Issued: **2016-02-04**

Date: **2020-05-05**

Expiry Date: **2024-05-26**

...making excellence a habit.™

Page 3 of 4

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.
This certificate was issued electronically and is bound by the conditions of the contract.

EC Design-Examination Certificate

Supplementary Information to CE 633673

Issued To: **Internacional Farmacéutica, S.A. de C.V.**
Carreteraco No. 44
Colonia Parque San Andrés
Coyoacán
C.P. 04040
Ciudad de México
Mexico

Certificate History

Date	Reference Number	Action
04 February 2016	10154728	First issue. Transfer from another Notified Body and renewal.
25 February 2019	8588105	Traceable to NB 0086.
Current	3061068	Certificate Renewal. Addition of a sterilization site, Proestéril, S.A de C.V. as a critical subcontractor for ETO sterilization for additional sterilization capacity Change of sterilization chamber from Chamber #4 to Chamber #1 at Sterigenics. Administrative address change, Legal Manufacturer Administrative address change, EU Rep Addition of needle codes in the 'supplementary information' table of certificate.

First Issued: **2016-02-04**

Date: **2020-05-05**

Expiry Date: **2024-05-26**

...making excellence a habit.™

Page 4 of 4

Validity of this certificate is conditional on the quality system being maintained to the requirements of the Directive as demonstrated through the required surveillance activities of the Notified Body.
 This certificate was issued electronically and is bound by the conditions of the contract.